
Midmark M11 - v1.0.0
Total Cycles: 5231

Name: RITTER 1
Sterilizer ID: V2178198
01/09/2025 01:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.7 F 0.0
2:00 98.3 F 0.0
4:00 120.5 F 0.0
6:00 147.1 F 0.2
8:00 168.9 F 2.8
10:00 189.3 F 5.9
12:00 203.9 F 9.1
14:00 221.9 F 12.9
16:00 224.1 F 7.4
18:00 236.0 F 12.0
20:00 234.4 F 8.8
22:00 240.7 F 10.7
24:00 245.0 F 12.8
26:00 256.8 F 19.0
28:00 267.3 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.1 F 28.7
2:00 272.4 F 29.1
3:00 271.4 F 28.3
4:00 272.1 F 28.7

Min 271.1 F 28.2
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 201.4 F 0.0
10:00 192.2 F 0.0
15:00 196.2 F 0.0
20:00 224.4 F 0.0
25:00 232.3 F 0.0
30:00 231.4 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 29:26
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:07:58
CYCLE COMPLETE

01/09/2025 02:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____