
Midmark M11 - v1.0.0
Total Cycles: 4314

Name: RITTER 1
Sterilizer ID: V2178198
02/21/2024 10:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.2 F 0.0
2:00 109.1 F 0.6
4:00 111.2 F 2.3
6:00 135.3 F 4.8
8:00 172.7 F 7.8
10:00 190.2 F 12.0
12:00 207.5 F 3.3
14:00 216.0 F 6.7
16:00 226.1 F 10.4
18:00 232.8 F 8.1
20:00 238.2 F 10.9
22:00 238.3 F 8.0
24:00 247.5 F 14.1
26:00 257.7 F 19.6
28:00 267.3 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.4 F 29.2
2:00 272.0 F 28.6
3:00 271.2 F 28.4
4:00 272.4 F 29.2

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 213.5 F 0.0
10:00 215.1 F 0.0
15:00 223.3 F 0.0
20:00 223.8 F 0.0
25:00 224.2 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 29:38
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:08:21
CYCLE COMPLETE

02/21/2024 11:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____