

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1094

Name: RITTER 1  
Sterilizer ID: V2178198  
08/18/2021 02:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.0 F 0.0  
2:00 151.1 F 1.2  
4:00 168.4 F 3.3  
6:00 193.2 F 6.3  
8:00 213.4 F 10.6  
10:00 217.0 F 6.4  
12:00 227.3 F 8.4  
14:00 237.8 F 12.8  
16:00 236.3 F 9.7  
18:00 238.4 F 9.6  
20:00 246.2 F 13.6  
22:00 257.3 F 19.5  
24:00 267.5 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.6 F 29.3  
2:00 271.8 F 28.7  
3:00 271.8 F 29.1  
4:00 272.6 F 29.3

Min 271.3 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.2 F 0.0  
5:00 222.1 F 0.0  
10:00 233.3 F 0.0  
15:00 225.1 F 0.0  
20:00 221.0 F 0.0  
25:00 223.8 F 0.0  
30:00 225.0 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 25:29  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:04:16  
CYCLE COMPLETE

08/18/2021 03:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_