
Midmark M11 - v1.0.5
Total Cycles: 2044

Name: RITTER2
Sterilizer ID: V2313228
02/14/2023 08:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.1 F 0.0
2:00 80.9 F 0.0
4:00 101.5 F 0.0
6:00 156.7 F 0.0
8:00 186.1 F 0.0
10:00 203.1 F 2.2
12:00 209.9 F 5.0
14:00 220.1 F 8.1
16:00 231.5 F 11.9
18:00 230.0 F 6.9
20:00 238.7 F 11.7
22:00 235.8 F 9.1
24:00 240.8 F 10.3
26:00 246.5 F 13.7
28:00 257.9 F 20.0
30:00 268.3 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.8
1:00 272.7 F 29.6
2:00 271.7 F 28.7
3:00 272.8 F 29.5
4:00 271.5 F 28.7

Min 271.4 F 28.6
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 205.5 F 0.0
10:00 212.5 F 0.0
15:00 211.6 F 0.0
20:00 209.7 F 0.0
25:00 217.3 F 0.0
30:00 213.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:08
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:06
CYCLE COMPLETE

02/14/2023 09:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____