
Midmark M11 - v1.0.5
Total Cycles: 3929

Name: RITTER2
Sterilizer ID: V2313228
09/11/2024 12:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.8 F 0.0
2:00 162.1 F 2.0
4:00 167.8 F 5.3
6:00 197.1 F 9.2
8:00 218.8 F 9.3
10:00 224.7 F 7.7
12:00 233.8 F 12.3
14:00 234.0 F 9.2
16:00 239.8 F 10.6
18:00 245.1 F 13.3
20:00 256.3 F 19.3
22:00 266.6 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.4 F 29.8
2:00 271.6 F 29.2
3:00 271.7 F 29.8
4:00 272.2 F 29.6

Min 271.1 F 29.0
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 220.6 F 0.0
10:00 233.1 F 0.0
15:00 228.8 F 0.0
20:00 227.4 F 0.0
25:00 226.0 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 23:40
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:02:27
CYCLE COMPLETE

09/11/2024 01:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____