
Midmark M11 - v1.0.0
Total Cycles: 3079

Name: RITTER 1
Sterilizer ID: V2178198
04/10/2023 08:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.5 F 0.0
2:00 171.3 F 1.5
4:00 176.5 F 4.6
6:00 201.6 F 9.4
8:00 220.0 F 7.0
10:00 229.8 F 9.1
12:00 237.1 F 10.1
14:00 242.3 F 12.0
16:00 240.7 F 10.6
18:00 254.1 F 17.3
20:00 265.8 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 272.2 F 28.6
2:00 272.6 F 29.1
3:00 271.6 F 28.2
4:00 272.7 F 29.0

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 219.6 F 0.0
10:00 213.8 F 0.0
15:00 219.1 F 0.0
20:00 229.9 F 0.0
25:00 227.6 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

04/10/2023 09:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____