

-----  
Midmark M11 - v1.0.5  
Total Cycles: 532

Name: RITTER2  
Sterilizer ID: V2313228  
11/10/2021 07:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.4 F 0.0  
2:00 77.6 F 0.0  
4:00 83.8 F 0.0  
6:00 95.3 F 0.1  
8:00 139.5 F 0.2  
10:00 170.0 F 0.3  
12:00 194.2 F 1.7  
14:00 210.9 F 4.9  
16:00 216.8 F 8.1  
18:00 227.0 F 11.8  
20:00 229.2 F 8.0  
22:00 239.3 F 13.0  
24:00 240.1 F 11.2  
26:00 238.9 F 10.1  
28:00 251.5 F 16.3  
30:00 262.2 F 22.5  
32:00 272.2 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.7  
1:00 272.1 F 29.0  
2:00 272.0 F 29.4  
3:00 271.3 F 28.6  
4:00 272.1 F 29.2  
  
Min 271.1 F 28.6  
Max 272.3 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.6 F 0.0  
5:00 214.9 F 0.0  
10:00 217.2 F 0.0  
15:00 225.8 F 0.0  
20:00 223.9 F 0.0  
25:00 207.9 F 0.0  
30:00 205.5 F 0.0  
  
Drying Complete

FILLING: 0:54  
HEATING: 32:27  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:10:29  
CYCLE COMPLETE

11/10/2021 08:15 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_