

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5003

Name: RITTER 1  
Sterilizer ID: V2178198  
10/17/2024 08:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.4 F 0.0  
2:00 72.0 F 0.0  
4:00 87.4 F 0.0  
6:00 112.3 F 0.1  
8:00 143.7 F 0.4  
10:00 170.6 F 2.0  
12:00 191.4 F 5.0  
14:00 217.1 F 8.0  
16:00 222.3 F 11.8  
18:00 224.7 F 5.3  
20:00 234.9 F 9.6  
22:00 237.5 F 10.2  
24:00 242.4 F 11.3  
26:00 237.7 F 9.0  
28:00 251.0 F 15.4  
30:00 262.1 F 21.7  
32:00 272.1 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 27.9  
1:00 272.6 F 28.6  
2:00 271.7 F 27.8  
3:00 272.0 F 28.5  
4:00 272.4 F 28.4  
  
Min 271.2 F 27.7  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.6 F 0.0  
5:00 194.8 F 0.0  
10:00 212.7 F 0.0  
15:00 221.1 F 0.0  
20:00 242.7 F 0.0  
25:00 234.1 F 0.0  
30:00 229.6 F 0.0  
  
Drying Complete

FILLING: 0:42  
HEATING: 32:33  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:10:41  
CYCLE COMPLETE

10/17/2024 09:35 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_