

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 002968  
16:35 13/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:38  
PRESSURIZING 2:42  
STERILIZING 3:37  
136.0 C 220kPa 3:37  
136.4 C 225kPa 4:07  
136.4 C 225kPa 4:37  
136.5 C 225kPa 5:07  
136.4 C 224kPa 5:37  
136.3 C 222kPa 6:07  
136.0 C 218kPa 6:37  
136.4 C 220kPa 7:07  
136.4 C 220kPa 7:07  
VENTING 7:07  
AIR DRYING 8:03  
CYCLE COMPLETE 8:10

-----  
Digital Signature #  
5BC5C1A088092E13  
-----