

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3382

Name: RITTER 1  
Sterilizer ID: V2178198  
06/13/2023 01:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.0 F 0.0  
2:00 187.6 F 2.2  
4:00 204.3 F 7.3  
6:00 224.4 F 13.1  
8:00 229.8 F 8.8  
10:00 239.6 F 10.5  
12:00 242.5 F 12.5  
14:00 241.6 F 11.4  
16:00 255.1 F 18.2  
18:00 266.7 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.6 F 29.2  
2:00 271.9 F 28.6  
3:00 271.2 F 28.4  
4:00 272.4 F 29.3  
  
Min 271.2 F 28.2  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.6 F 0.0  
5:00 214.8 F 0.0  
10:00 225.3 F 0.0  
15:00 224.1 F 0.0  
20:00 224.4 F 0.0  
25:00 224.6 F 0.0  
30:00 223.2 F 0.0  
  
Drying Complete

FILLING: 1:48  
HEATING: 19:33  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:58:41  
CYCLE COMPLETE

06/13/2023 02:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_