

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000039  
12:21 19/02/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:48  
PRESSURIZING 1:54  
STERILIZING 2:58  
136.3 C 224kPa 2:58  
136.2 C 224kPa 3:28  
136.0 C 223kPa 3:58  
136.2 C 224kPa 4:28  
136.3 C 225kPa 4:58  
136.5 C 227kPa 5:28  
136.0 C 221kPa 5:58  
136.3 C 225kPa 6:28  
136.3 C 225kPa 6:28  
VENTING 6:28  
AIR DRYING 7:12  
CYCLE COMPLETE 22:13  
-----

Digital Signature #  
6DD44FD2629F8250  
-----