
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004214
14:56 16/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:13
PRESSURIZING 2:16
STERILIZING 3:13
136.0 C 220kPa 3:13
136.3 C 225kPa 3:43
136.4 C 225kPa 4:13
136.4 C 225kPa 4:43
136.3 C 224kPa 5:13
136.4 C 224kPa 5:43
136.4 C 225kPa 6:13
136.4 C 221kPa 6:43
136.4 C 221kPa 6:43
VENTING 6:43
AIR DRYING 7:32
CYCLE COMPLETE 8:14

Digital Signature #
C0CD480103CFE148
