
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002409
9:00 10/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:39
STERILIZING 3:47
136.0 C 220kPa 3:47
136.0 C 220kPa 4:17
136.3 C 223kPa 4:47
136.4 C 223kPa 5:17
136.2 C 223kPa 5:47
136.3 C 221kPa 6:17
136.2 C 222kPa 6:47
136.3 C 223kPa 7:17
136.3 C 223kPa 7:17
VENTING 7:17
AIR DRYING 8:08
CYCLE COMPLETE 38:08

Digital Signature #
FC100D9EC5E8CC6A
