
Midmark M11 - v1.0.0
Total Cycles: 4780

Name: RITTER 1
Sterilizer ID: V2178198
08/01/2024 08:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 74.6 F 0.0
4:00 103.4 F 0.2
6:00 120.9 F 1.6
8:00 148.7 F 3.8
10:00 171.1 F 6.2
12:00 187.2 F 8.8
14:00 198.8 F 12.1
16:00 207.3 F 3.3
18:00 218.6 F 6.6
20:00 228.4 F 10.2
22:00 233.2 F 9.4
24:00 237.1 F 10.2
26:00 237.6 F 9.1
28:00 246.0 F 13.2
30:00 256.4 F 18.7
32:00 266.1 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.0
2:00 272.1 F 28.5
3:00 271.5 F 28.3
4:00 272.6 F 29.1

Min 271.5 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 184.9 F 0.0
10:00 186.5 F 0.0
15:00 199.4 F 0.0
20:00 207.8 F 0.0
25:00 213.3 F 0.0
30:00 214.8 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 33:52
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:12:16
CYCLE COMPLETE

08/01/2024 09:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____