
Midmark M11 - v1.0.0
Total Cycles: 1517

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2022 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.9 F 0.0
2:00 205.4 F 4.3
4:00 217.1 F 9.9
6:00 222.6 F 6.0
8:00 236.4 F 11.0
10:00 235.7 F 8.5
12:00 240.8 F 10.3
14:00 247.8 F 14.0
16:00 260.0 F 20.9
18:00 271.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 28.8
2:00 271.7 F 28.5
3:00 272.5 F 28.9
4:00 271.5 F 28.3

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.8 F 0.0
5:00 236.7 F 0.0
10:00 228.9 F 0.0
15:00 227.3 F 0.0
20:00 230.9 F 0.0
25:00 232.9 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 2:35
HEATING: 18:39
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:25
CYCLE COMPLETE

01/12/2022 03:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____