
Midmark M11 - v1.0.5
Total Cycles: 2539

Name: RITTER2
Sterilizer ID: V2313228
06/20/2023 12:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.8 F 0.0
2:00 193.7 F 3.9
4:00 209.0 F 9.4
6:00 224.2 F 7.6
8:00 234.3 F 11.2
10:00 235.0 F 9.0
12:00 238.8 F 9.8
14:00 248.2 F 14.7
16:00 260.4 F 21.5
18:00 271.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.5 F 29.7
2:00 272.3 F 29.1
3:00 271.2 F 28.7
4:00 272.2 F 29.7

Min 271.1 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 231.8 F 0.0
10:00 226.3 F 0.0
15:00 222.5 F 0.0
20:00 221.4 F 0.0
25:00 222.8 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 2:23
HEATING: 18:45
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

06/20/2023 01:41 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____