
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001948
11:45 12/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:40
STERILIZING 3:44
136.0 C 220kPa 3:44
136.2 C 225kPa 4:14
136.4 C 225kPa 4:44
136.4 C 223kPa 5:14
136.4 C 226kPa 5:44
136.4 C 226kPa 6:14
136.4 C 226kPa 6:44
136.4 C 226kPa 7:14
136.4 C 226kPa 7:14
VENTING 7:14
AIR DRYING 8:13
CYCLE COMPLETE 17:53

Digital Signature #
2602AE4C7B7F4C5D
