
Midmark M11 - v1.0.0
Total Cycles: 3496

Name: RITTER 1
Sterilizer ID: V2178198
07/06/2023 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.2 F 0.0
2:00 197.4 F 2.6
4:00 222.6 F 8.0
6:00 234.1 F 12.3
8:00 234.2 F 9.9
10:00 238.0 F 8.4
12:00 244.0 F 12.1
14:00 245.7 F 13.1
16:00 258.6 F 20.0
18:00 270.0 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 272.5 F 29.1
2:00 271.5 F 28.3
3:00 272.5 F 29.0
4:00 271.5 F 28.3

Min 271.5 F 28.3
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 228.8 F 0.0
10:00 227.7 F 0.0
15:00 229.1 F 0.0
20:00 228.9 F 0.0
25:00 229.0 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 2:07
HEATING: 18:46
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

07/06/2023 01:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____