

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001531  
14:25 23/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:22  
PRESSURIZING 2:27  
STERILIZING 3:27  
136.4 C 224kPa 3:27  
136.3 C 224kPa 3:57  
136.4 C 223kPa 4:27  
136.4 C 224kPa 4:57  
136.2 C 224kPa 5:27  
136.4 C 226kPa 5:57  
136.4 C 225kPa 6:27  
136.4 C 224kPa 6:57  
136.4 C 224kPa 6:57  
VENTING 6:57  
AIR DRYING 7:53  
CYCLE COMPLETE 7:58

-----  
Digital Signature #  
98D652BED74A9713  
-----