
Midmark M11 - v1.0.0
Total Cycles: 893

Name: RITTER 1
Sterilizer ID: V2178198
06/14/2021 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.6 F 0.0
2:00 165.1 F 1.3
4:00 161.5 F 4.0
6:00 179.9 F 7.6
8:00 211.9 F 12.2
10:00 216.7 F 5.7
12:00 228.5 F 10.4
14:00 232.6 F 8.0
16:00 242.9 F 12.7
18:00 241.2 F 11.1
20:00 253.8 F 17.5
22:00 265.0 F 24.1
24:00 271.2 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.5 F 29.2
2:00 272.2 F 29.5
3:00 272.1 F 29.0
4:00 272.6 F 29.6

Min 271.6 F 28.8
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.6 F 0.0
5:00 243.1 F 0.0
10:00 229.9 F 0.0
15:00 225.2 F 0.0
20:00 224.1 F 0.0
25:00 230.4 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:38
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:03:01
CYCLE COMPLETE

06/14/2021 01:26 PM
ACCEPTED: X
REJECTED:
Operator: U01