
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001700
12:44 07/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:27
STERILIZING 3:26
136.0 C 220kPa 3:26
136.4 C 225kPa 3:56
136.4 C 225kPa 4:26
136.3 C 225kPa 4:56
136.4 C 223kPa 5:26
136.4 C 225kPa 5:56
136.3 C 224kPa 6:26
136.4 C 225kPa 6:56
136.4 C 225kPa 6:56
VENTING 6:56
AIR DRYING 7:47
CYCLE COMPLETE 10:17

Digital Signature #
179314DFAAD905C1
