
Midmark M11 - v1.0.0
Total Cycles: 3584

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2023 09:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.3 F 0.0
2:00 181.8 F 1.4
4:00 184.0 F 5.0
6:00 203.9 F 9.2
8:00 219.0 F 11.3
10:00 224.4 F 7.1
12:00 236.2 F 11.9
14:00 235.3 F 8.9
16:00 240.8 F 10.5
18:00 246.3 F 13.3
20:00 257.9 F 19.5
22:00 268.4 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 28.8
2:00 271.5 F 28.2
3:00 272.7 F 28.9
4:00 271.8 F 28.2

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 196.0 F 0.0
10:00 200.3 F 0.0
15:00 198.5 F 0.0
20:00 195.3 F 0.0
25:00 195.2 F 0.0
30:00 198.2 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 23:15
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:01:53
CYCLE COMPLETE

07/26/2023 10:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____