

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2405

Name: RITTER 1  
Sterilizer ID: V2178198  
10/05/2022 03:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.4 F 0.0  
2:00 157.8 F 2.3  
4:00 175.7 F 5.7  
6:00 207.1 F 10.3  
8:00 218.4 F 4.9  
10:00 232.4 F 9.7  
12:00 236.5 F 9.4  
14:00 241.3 F 11.8  
16:00 239.0 F 10.0  
18:00 252.0 F 16.5  
20:00 263.5 F 23.1  
22:00 272.0 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.4  
1:00 272.4 F 29.1  
2:00 271.3 F 28.3  
3:00 272.4 F 29.2  
4:00 271.8 F 28.6

Min 271.1 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.2 F 0.0  
5:00 234.2 F 0.0  
10:00 235.5 F 0.0  
15:00 229.7 F 0.0  
20:00 229.2 F 0.0  
25:00 229.1 F 0.0  
30:00 224.5 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 22:12  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:24  
CYCLE COMPLETE

10/05/2022 04:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_