
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000588
15:01 16/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:29
STERILIZING 3:36
136.0 C 228kPa 3:36
136.1 C 231kPa 4:06
136.4 C 233kPa 4:36
136.4 C 232kPa 5:06
136.1 C 233kPa 5:36
136.3 C 232kPa 6:06
136.1 C 232kPa 6:36
136.3 C 232kPa 7:06
136.3 C 232kPa 7:06
VENTING 7:06
AIR DRYING 7:56
CYCLE COMPLETE 22:56

Digital Signature #
9A743E8DE8826F6A
