
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001152
10:43 05/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:53
STERILIZING 4:01
136.4 C 222kPa 4:01
136.4 C 225kPa 4:31
136.0 C 221kPa 5:01
136.2 C 223kPa 5:31
136.4 C 226kPa 6:01
136.4 C 222kPa 6:31
136.4 C 225kPa 7:01
136.4 C 225kPa 7:31
136.4 C 225kPa 7:31
VENTING 7:31
AIR DRYING 8:23
CYCLE COMPLETE 14:45

Digital Signature #
ACBF79EF1023A41E
