
Midmark M11 - v1.0.5
Total Cycles: 4162

Name: RITTER2
Sterilizer ID: V2313228
11/27/2024 03:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.2 F 0.0
2:00 87.0 F 0.0
4:00 90.2 F 0.0
6:00 151.8 F 0.0
8:00 184.2 F 0.2
10:00 205.8 F 0.6
12:00 213.9 F 3.5
14:00 224.6 F 6.8
16:00 233.1 F 10.5
18:00 234.9 F 8.9
20:00 242.1 F 12.3
22:00 240.0 F 10.9
24:00 238.3 F 9.9
26:00 251.7 F 16.4
28:00 263.0 F 23.0
30:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.6 F 29.3
2:00 271.4 F 28.5
3:00 272.5 F 29.5
4:00 272.1 F 28.9

Min 271.0 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.2 F 0.0
5:00 233.9 F 0.0
10:00 220.7 F 0.0
15:00 182.3 F 0.0
20:00 179.4 F 0.0
25:00 178.7 F 0.0
30:00 179.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 30:16
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:08:57
CYCLE COMPLETE

11/27/2024 05:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____