

-----  
Midmark M11 - v1.0.5  
Total Cycles: 911

Name: RITTER2  
Sterilizer ID: V2313228  
03/28/2022 07:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.1 F 0.0  
2:00 76.7 F 0.0  
4:00 100.1 F 0.0  
6:00 112.7 F 0.5  
8:00 150.5 F 1.5  
10:00 188.2 F 1.9  
12:00 211.0 F 4.0  
14:00 224.4 F 7.4  
16:00 228.0 F 11.2  
18:00 230.3 F 7.3  
20:00 241.2 F 12.9  
22:00 241.0 F 11.2  
24:00 239.4 F 10.3  
26:00 253.3 F 17.2  
28:00 264.8 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.5 F 29.5  
2:00 272.0 F 29.0  
3:00 271.0 F 28.6  
4:00 272.3 F 29.6  
  
Min 271.0 F 28.5  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.0 F 0.0  
5:00 215.5 F 0.0  
10:00 206.8 F 0.0  
15:00 208.7 F 0.0  
20:00 213.2 F 0.0  
25:00 221.3 F 0.0  
30:00 219.0 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 29:55  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:07:51  
CYCLE COMPLETE

03/28/2022 08:08 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_