
Midmark M11 - v1.0.0
Total Cycles: 4150

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2023 04:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.4 F 0.0
2:00 190.1 F 4.3
4:00 197.7 F 9.7
6:00 218.5 F 6.2
8:00 232.4 F 9.4
10:00 238.3 F 8.7
12:00 244.0 F 12.7
14:00 242.9 F 11.3
16:00 256.6 F 18.4
18:00 268.1 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 272.3 F 28.3
2:00 272.5 F 28.6
3:00 271.9 F 28.0
4:00 272.8 F 28.7

Min 271.6 F 27.8
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 224.5 F 0.0
10:00 239.9 F 0.0
15:00 227.8 F 0.0
20:00 222.9 F 0.0
25:00 222.9 F 0.0
30:00 220.9 F 0.0

Drying Complete

FILLING: 1:54
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

12/21/2023 05:45 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____