
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004734
11:56 21/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:00
STERILIZING 3:05
136.4 C 220kPa 3:05
136.3 C 223kPa 3:35
136.4 C 224kPa 4:05
136.4 C 223kPa 4:35
136.3 C 223kPa 5:05
136.3 C 224kPa 5:35
136.2 C 223kPa 6:05
136.4 C 225kPa 6:35
136.4 C 225kPa 6:35
VENTING 6:35
AIR DRYING 7:25
CYCLE COMPLETE 7:29

Digital Signature #
FF5E5E0E9A634D55
