
Midmark M11 - v1.0.5
Total Cycles: 1097

Name: RITTER2
Sterilizer ID: V2313228
05/11/2022 02:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.6 F 0.0
2:00 194.3 F 4.5
4:00 207.9 F 10.1
6:00 223.7 F 6.1
8:00 234.7 F 11.7
10:00 237.2 F 9.9
12:00 237.3 F 8.9
14:00 251.0 F 16.0
16:00 263.1 F 23.1
18:00 272.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.3 F 29.3
2:00 271.4 F 28.8
3:00 272.4 F 29.4
4:00 271.7 F 28.7

Min 271.3 F 28.6
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 224.3 F 0.0
10:00 233.5 F 0.0
15:00 227.2 F 0.0
20:00 223.5 F 0.0
25:00 222.2 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 2:37
HEATING: 18:10
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

05/11/2022 03:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____