
Midmark M11 - v1.0.5
Total Cycles: 1538

Name: RITTER2
Sterilizer ID: V2313228
09/26/2022 04:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.5 F 0.0
2:00 197.1 F 4.2
4:00 202.4 F 9.6
6:00 219.4 F 7.5
8:00 232.6 F 11.9
10:00 234.6 F 10.0
12:00 237.0 F 8.7
14:00 249.5 F 16.0
16:00 261.8 F 23.0
18:00 272.8 F 30.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.5
1:00 272.4 F 30.3
2:00 271.7 F 29.4
3:00 272.3 F 30.2
4:00 271.9 F 29.6

Min 271.5 F 29.4
Max 272.4 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 220.0 F 0.0
10:00 226.6 F 0.0
15:00 233.0 F 0.0
20:00 227.2 F 0.0
25:00 227.0 F 0.0
30:00 227.9 F 0.0

Drying Complete

FILLING: 2:26
HEATING: 18:13
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

09/26/2022 05:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____