
Midmark M11 - v1.0.0
Total Cycles: 4804

Name: RITTER 1
Sterilizer ID: V2178198
08/12/2024 12:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.0 F 0.0
2:00 160.5 F 1.7
4:00 185.4 F 5.0
6:00 204.8 F 8.8
8:00 215.8 F 13.3
10:00 221.4 F 7.1
12:00 232.1 F 11.5
14:00 232.5 F 8.0
16:00 245.1 F 13.5
18:00 243.1 F 11.9
20:00 254.5 F 17.8
22:00 264.6 F 23.9
24:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 271.6 F 28.7
2:00 272.1 F 29.1
3:00 272.5 F 29.4
4:00 272.7 F 29.4

Min 271.2 F 28.3
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 216.1 F 0.0
10:00 213.4 F 0.0
15:00 211.0 F 0.0
20:00 207.6 F 0.0
25:00 208.6 F 0.0
30:00 212.2 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 24:19
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:03:28
CYCLE COMPLETE

08/12/2024 01:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____