

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000166  
7:32 16/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:32  
PRESSURIZING 4:32  
STERILIZING 5:45  
136.0 C 220kPa 5:45  
136.3 C 225kPa 6:15  
136.4 C 225kPa 6:45  
136.4 C 223kPa 7:15  
136.2 C 223kPa 7:45  
136.4 C 223kPa 8:15  
136.4 C 225kPa 8:45  
136.3 C 225kPa 9:15  
136.3 C 225kPa 9:15  
VENTING 9:15  
AIR DRYING 10:16  
CYCLE COMPLETE 25:16  
-----

Digital Signature #  
2BBD5FADA2F09E7A  
-----