
Midmark M11 - v1.0.5
Total Cycles: 1162

Name: RITTER2
Sterilizer ID: V2313228
06/06/2022 03:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.9 F 0.0
2:00 180.5 F 3.6
4:00 187.1 F 8.4
6:00 210.2 F 10.1
8:00 223.4 F 8.9
10:00 234.4 F 10.2
12:00 241.4 F 12.4
14:00 241.0 F 11.3
16:00 254.1 F 18.0
18:00 265.2 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.6
2:00 271.4 F 28.9
3:00 272.5 F 29.7
4:00 271.9 F 29.2

Min 271.3 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 230.9 F 0.0
10:00 230.8 F 0.0
15:00 230.7 F 0.0
20:00 227.1 F 0.0
25:00 230.4 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:30
CYCLE COMPLETE

06/06/2022 04:45 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____