
Midmark M11 - v1.0.0
Total Cycles: 1913

Name: RITTER 1
Sterilizer ID: V2178198
05/05/2022 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.2 F 0.0
2:00 191.5 F 3.3
4:00 189.0 F 7.6
6:00 210.4 F 12.9
8:00 229.6 F 7.5
10:00 239.7 F 12.6
12:00 238.2 F 9.5
14:00 237.2 F 8.8
16:00 249.5 F 14.6
18:00 261.8 F 21.6
20:00 272.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 27.9
1:00 272.6 F 28.6
2:00 271.2 F 27.8
3:00 272.4 F 28.7
4:00 272.1 F 28.2

Min 271.0 F 27.8
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 238.7 F 0.0
10:00 230.1 F 0.0
15:00 227.5 F 0.0
20:00 228.5 F 0.0
25:00 229.3 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

05/05/2022 11:13 AM
ACCEPTED: X
REJECTED:
Operator: U02