

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2225

Name: RITTER 1  
Sterilizer ID: V2178198  
08/16/2022 06:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.1 F 0.0  
2:00 74.8 F 0.0  
4:00 77.2 F 0.0  
6:00 94.3 F 0.6  
8:00 131.4 F 3.3  
10:00 160.3 F 6.1  
12:00 219.8 F 10.1  
14:00 227.9 F 6.9  
16:00 225.0 F 7.0  
18:00 234.3 F 10.7  
20:00 234.5 F 7.8  
22:00 239.7 F 11.0  
24:00 233.8 F 7.6  
26:00 247.8 F 13.9  
28:00 258.6 F 19.8  
30:00 268.5 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 271.3 F 28.8  
2:00 271.7 F 29.0  
3:00 271.8 F 29.1  
4:00 271.8 F 29.1  
  
Min 271.2 F 28.6  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.5 F 0.0  
5:00 221.7 F 0.0  
10:00 196.2 F 0.0  
15:00 230.6 F 0.0  
20:00 224.3 F 0.0  
25:00 230.4 F 0.0  
30:00 227.1 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:55  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:09:54  
CYCLE COMPLETE

08/16/2022 08:08 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_