
Midmark M11 - v1.0.0
Total Cycles: 5011

Name: RITTER 1
Sterilizer ID: V2178198
10/21/2024 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.9 F 0.0
2:00 72.7 F 0.0
4:00 107.7 F 0.0
6:00 125.3 F 0.1
8:00 154.6 F 0.2
10:00 179.4 F 0.4
12:00 199.8 F 2.6
14:00 206.7 F 5.7
16:00 216.6 F 8.7
18:00 227.0 F 12.2
20:00 227.7 F 6.7
22:00 238.8 F 11.3
24:00 234.4 F 8.0
26:00 244.4 F 11.8
28:00 245.1 F 12.5
30:00 256.9 F 18.7
32:00 267.4 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.6 F 28.4
2:00 271.6 F 27.9
3:00 272.7 F 28.5
4:00 271.6 F 27.8

Min 271.5 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.6 F 0.0
5:00 182.8 F 0.0
10:00 195.2 F 0.0
15:00 191.2 F 0.0
20:00 191.3 F 0.0
25:00 195.6 F 0.0
30:00 208.5 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 33:27
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:11:32
CYCLE COMPLETE

10/21/2024 11:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____