
Midmark M11 - v1.0.5
Total Cycles: 99

Name: RITTER2
Sterilizer ID: V2313228
06/15/2021 06:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.1 F 0.0
2:00 77.3 F 0.2
4:00 94.6 F 0.8
6:00 97.1 F 2.7
8:00 131.5 F 5.4
10:00 164.2 F 8.5
12:00 190.3 F 12.1
14:00 215.9 F 4.8
16:00 225.4 F 8.3
18:00 234.6 F 12.1
20:00 231.2 F 8.0
22:00 242.2 F 12.9
24:00 240.3 F 11.1
26:00 252.1 F 16.9
28:00 262.2 F 22.9
30:00 272.1 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 271.9 F 29.8
2:00 272.5 F 30.0
3:00 272.5 F 29.8
4:00 272.3 F 29.8

Min 271.5 F 29.4
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 234.7 F 0.0
10:00 230.9 F 0.0
15:00 223.2 F 0.0
20:00 219.0 F 0.0
25:00 217.1 F 0.0
30:00 216.9 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 31:02
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:19
CYCLE COMPLETE

06/15/2021 07:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____