
Midmark M11 - v1.0.0
Total Cycles: 2460

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2022 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.5 F 0.0
2:00 182.3 F 3.9
4:00 197.7 F 8.4
6:00 212.4 F 13.2
8:00 230.9 F 8.4
10:00 240.1 F 10.9
12:00 241.1 F 11.2
14:00 237.4 F 9.2
16:00 251.2 F 15.9
18:00 263.1 F 22.6
20:00 272.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 272.0 F 29.2
2:00 272.1 F 29.3
3:00 272.1 F 29.2
4:00 272.1 F 29.2

Min 271.8 F 28.8
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 222.2 F 0.0
10:00 232.7 F 0.0
15:00 227.8 F 0.0
20:00 226.1 F 0.0
25:00 230.0 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 20:54
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:28
CYCLE COMPLETE

10/19/2022 03:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____