

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 002338  
15:26 29/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 1:57  
STERILIZING 2:53  
136.0 C 220kPa 2:53  
136.4 C 225kPa 3:23  
136.4 C 224kPa 3:53  
136.4 C 223kPa 4:23  
136.1 C 222kPa 4:53  
136.4 C 225kPa 5:23  
136.3 C 223kPa 5:53  
136.4 C 225kPa 6:23  
136.4 C 225kPa 6:23  
VENTING 6:23  
AIR DRYING 7:13  
CYCLE COMPLETE 10:31  
-----

Digital Signature #  
6737EE299E21F7CD  
-----