
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000699
15:47 22/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:22
STERILIZING 3:24
136.4 C 222kPa 3:24
136.4 C 224kPa 3:54
136.4 C 221kPa 4:24
136.4 C 225kPa 4:54
136.4 C 222kPa 5:24
136.4 C 223kPa 5:54
136.2 C 223kPa 6:24
136.4 C 224kPa 6:54
136.4 C 224kPa 6:54
VENTING 6:55
AIR DRYING 7:47
CYCLE COMPLETE 7:51

Digital Signature #
15206BB8D7326073
