
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002366
12:18 06/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:48
PRESSURIZING 1:58
STERILIZING 2:53
136.0 C 221kPa 2:53
136.4 C 222kPa 3:23
136.2 C 225kPa 3:53
136.2 C 224kPa 4:23
136.4 C 226kPa 4:53
136.1 C 222kPa 5:23
136.4 C 223kPa 5:53
136.1 C 223kPa 6:23
136.1 C 223kPa 6:23
VENTING 6:23
AIR DRYING 7:13
CYCLE COMPLETE 22:06

Digital Signature #
198F3867CBD875FE
