
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005234
11:53 15/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:25
STERILIZING 3:22
136.4 C 222kPa 3:22
136.4 C 225kPa 3:52
136.4 C 225kPa 4:22
136.0 C 222kPa 4:52
136.4 C 224kPa 5:22
136.4 C 225kPa 5:52
136.4 C 222kPa 6:22
136.4 C 225kPa 6:52
136.4 C 225kPa 6:52
VENTING 6:52
AIR DRYING 8:58
CYCLE COMPLETE 10:46

Digital Signature #
BCF8C395CFEB3061
