
Midmark M11 - v1.0.0
Total Cycles: 823

Name: RITTER 1
Sterilizer ID: V2178198
05/19/2021 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.9 F 0.0
2:00 173.2 F 3.1
4:00 188.8 F 7.4
6:00 209.3 F 12.6
8:00 222.5 F 6.5
10:00 235.9 F 11.8
12:00 235.4 F 8.8
14:00 240.1 F 9.9
16:00 247.5 F 13.8
18:00 260.0 F 20.6
20:00 271.1 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 28.5
2:00 271.6 F 28.3
3:00 272.5 F 28.6
4:00 271.5 F 28.1

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 233.8 F 0.0
10:00 238.9 F 0.0
15:00 223.7 F 0.0
20:00 218.8 F 0.0
25:00 222.7 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

05/19/2021 01:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____