

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5207

Name: RITTER 1  
Sterilizer ID: V2178198  
12/19/2024 10:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.6 F 0.0  
2:00 186.4 F 0.2  
4:00 187.9 F 1.9  
6:00 207.3 F 7.0  
8:00 223.0 F 11.9  
10:00 228.3 F 7.1  
12:00 239.9 F 12.8  
14:00 239.8 F 10.5  
16:00 237.0 F 8.9  
18:00 251.8 F 15.9  
20:00 263.6 F 22.7  
22:00 272.0 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.5 F 28.7  
2:00 271.7 F 27.9  
3:00 272.2 F 28.6  
4:00 272.4 F 28.5

Min 271.4 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.1 F 0.0  
5:00 212.5 F 0.0  
10:00 224.4 F 0.0  
15:00 237.6 F 0.0  
20:00 228.3 F 0.0  
25:00 229.8 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 22:10  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:00:35  
CYCLE COMPLETE

12/19/2024 11:53 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_