

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2121

Name: RITTER2  
Sterilizer ID: V2313228  
03/06/2023 11:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.6 F 0.0  
2:00 194.3 F 4.0  
4:00 203.3 F 9.2  
6:00 219.0 F 7.5  
8:00 231.5 F 10.6  
10:00 233.5 F 8.3  
12:00 239.8 F 10.5  
14:00 247.7 F 14.5  
16:00 260.5 F 21.7  
18:00 272.3 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.6 F 29.6  
2:00 272.0 F 29.0  
3:00 271.2 F 28.8  
4:00 272.6 F 29.7  
  
Min 271.1 F 28.6  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.1 F 0.0  
5:00 223.8 F 0.0  
10:00 229.7 F 0.0  
15:00 224.3 F 0.0  
20:00 225.0 F 0.0  
25:00 229.6 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 1:56  
HEATING: 18:32  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:37  
CYCLE COMPLETE

03/06/2023 12:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_