

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4501

Name: RITTER 1  
Sterilizer ID: V2178198  
04/29/2024 10:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.8 F 0.0  
2:00 163.2 F 2.1  
4:00 166.0 F 5.5  
6:00 190.0 F 9.5  
8:00 213.0 F 6.4  
10:00 219.4 F 7.9  
12:00 231.9 F 12.8  
14:00 233.6 F 9.3  
16:00 238.2 F 9.7  
18:00 245.4 F 13.2  
20:00 257.1 F 19.4  
22:00 267.7 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.7 F 29.2  
2:00 271.5 F 28.5  
3:00 272.2 F 29.4  
4:00 272.3 F 29.0

Min 271.3 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 219.4 F 0.0  
10:00 213.2 F 0.0  
15:00 213.8 F 0.0  
20:00 230.0 F 0.0  
25:00 222.9 F 0.0  
30:00 223.8 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 23:25  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:02:01  
CYCLE COMPLETE

04/29/2024 11:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_