
Midmark M11 - v1.0.0
Total Cycles: 5256

Name: RITTER 1
Sterilizer ID: V2178198
01/21/2025 01:54 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.1 F 0.0
2:00 80.0 F 0.0
4:00 93.3 F 1.5
6:00 132.5 F 4.4
8:00 160.0 F 7.7
10:00 179.5 F 11.9
12:00 198.6 F 2.8
14:00 214.1 F 6.9
16:00 226.1 F 10.8
18:00 226.8 F 6.4
20:00 239.3 F 11.2
22:00 235.6 F 8.4
24:00 250.6 F 15.5
26:00 262.2 F 22.1
28:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.4 F 28.5
2:00 272.9 F 29.0
3:00 271.9 F 28.4

Min 271.6 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 161.5 F 0.0
10:00 167.0 F 0.0
15:00 171.9 F 0.0
20:00 176.2 F 0.0
25:00 180.7 F 0.0
30:00 178.7 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 28:22
STERILIZING: 3:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:05:16
CYCLE COMPLETE

01/21/2025 02:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____