
Midmark M11 - v1.0.0
Total Cycles: 3673

Name: RITTER 1
Sterilizer ID: V2178198
08/17/2023 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.5 F 0.0
2:00 202.9 F 4.2
4:00 217.6 F 9.5
6:00 221.9 F 6.5
8:00 234.3 F 10.5
10:00 236.0 F 7.5
12:00 242.2 F 11.4
14:00 245.7 F 13.2
16:00 258.5 F 20.0
18:00 269.6 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.2 F 28.7
2:00 272.6 F 29.2
3:00 271.6 F 28.3
4:00 272.6 F 29.0

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 227.5 F 0.0
10:00 232.6 F 0.0
15:00 236.8 F 0.0
20:00 221.3 F 0.0
25:00 216.9 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 2:34
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

08/17/2023 03:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____