

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3903

Name: RITTER2  
Sterilizer ID: V2313228  
08/29/2024 10:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.0 F 0.0  
2:00 124.9 F 0.1  
4:00 136.2 F 0.4  
6:00 160.8 F 2.5  
8:00 180.4 F 6.0  
10:00 202.3 F 9.7  
12:00 216.5 F 10.0  
14:00 222.6 F 7.6  
16:00 228.1 F 11.9  
18:00 231.0 F 8.3  
20:00 242.3 F 13.5  
22:00 242.2 F 12.0  
24:00 253.8 F 17.8  
26:00 264.1 F 24.1  
28:00 271.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 272.4 F 29.8  
2:00 271.3 F 29.4  
3:00 272.3 F 29.7  
4:00 271.3 F 29.2

Min 271.2 F 29.1  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
5:00 213.7 F 0.0  
10:00 216.1 F 0.0  
15:00 230.2 F 0.0  
20:00 227.4 F 0.0  
25:00 233.0 F 0.0  
30:00 229.6 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 28:07  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:06:36  
CYCLE COMPLETE

08/29/2024 11:53 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_